

Contract notice: JO S 236-359198

Invitation to tender No. ENER/D4/250-1-2010 concerning

" Monitor for radioactive contamination measurements of large objects "

Time-limit for receipt of tenders: 25/02/2011

FREQUENTLY ASKED QUESTIONS

Version 14/02/2011

Question 1: Would a tender offering a comparable chamber volume but alternative chamber dimension be considered, i.e. greater height but less depth?

Answer 1: **The dimensions of the measurement chamber given by the technical specifications shall allow for radioactive contamination detection of specific large objects. During the evaluation procedure of the offers the proposed contamination monitors will be evaluated with a view whether those large objects fit into the measurement chamber.**

Question 2: What is the MDA requirement and are there any specific radio nuclides that the system is required to detect?

Answer 2: **The minimum detection levels for different isotopes and the related measurement time is an award criteria and shall be included in the technical section of the offer.**

Question 3: Is a search function within the measurement results database required or is the ability to view the saved measurement reports acceptable?

Answer 3: **Both options are acceptable.**

Question 4: Would a tender proposing stage payments be considered?

Answer 4: **A payment of the invoice will be made only following the acceptance of the final product by the European Commission.**

Question 5: Can you confirm that the secondary screen is for display only and is there a required distance from the system for this display screen.

Answer 5: **The secondary screen should show the same information as the primary screen. There are no minimum or maximum distance requirements between the measurement system and this screen.**