

Notification of a Body in the framework of a technical harmonization directive

From : Ministry of Economic
Development, Labor and
Technology
Plac Trzech Krzyzy 3/5
00-507 Warszawa
Poland

To : **European Commission**
GROWTH Directorate-General
200 Rue de la Loi,
B-1049 Brussels.
Other Member States

Reference :

Legislation : 2014/32/EU Measuring Instruments Directive

Body name, address, telephone, fax, email, website :

OKREGOWY URZAD MIAR W POZNANIU
ul. Krakowska 19
61-893 POZNAN
Poland
Phone : +48 61 856 72 40
Fax : +48 61 855 22 02
Email : oum.poznan@gum.gov.pl
Website :

Body :

NB 1441

The body is assessed according to :

EN ISO/IEC 17020 - Inspection

The competence of the body was assessed by : Minister of Economy and Finance

The assessment of the body covers the product categories and conformity assessment procedures concerned by this notification : Yes

Tasks performed by the Body :

Last approval date : 24/01/2017

Product family, product /Intended use/Product range	Procedure/Modules	Annexes or articles of the directives
Measuring Instruments for Liquids Other than Water (Annex VII MI-005) <ul style="list-style-type: none">- Measuring systems on pipeline (accuracy class 0,3)- All measuring systems (accuracy class 0,5), in particular<ul style="list-style-type: none">- fuel dispensers (not for liquefied gases)- measuring systems on road tankers for liquids of low viscosity (<20 mPa.s)- measuring systems for (un)loading ships and rail and road tankers- measuring systems for milk- Measuring systems for liquefied gases under pressure measured at a temperature equal to or above -10* (accuracy class 1,0)<ul style="list-style-type: none">- measuring systems normally in class 0,3 or 0,5 but used for liquids whose temperature is less than -10°C or greater than 50°C- measuring systems normally in class 0,3 or 0,5 but used for liquids whose dynamic viscosity is higher than 1 000 mPa.s- measuring systems normally in class 0,3 or 0,5 but used for liquids whose maximum volumetric flowrate is not higher than 20 L/h	Conformity to type based on instrument verification	Annex II - Module F
Automatic Weighing Instruments (Annex VIII MI-006)	Conformity to type based on instrument verification Conformity based on instrument verification	Annex II - Module F Annex II - Module F1