



JRC CRL - FCM Database / GC-MS

Sample Name: 2,4,6-Triamino-1,3,5-triazine
Solvent: Ethanol
Concentration: 100 µg/mL

Date: 13/09/2005
CAS: 000108-78-1
PM ref: 25420 [U.R.N. M001]



m/z	Abundance	Ion Intensity %	m/z	Abundance	Ion Intensity %
41.20	1097,0	4,40	67.10	459,0	1,84
42.10	6396,0	25,67	68.00	4334,0	17,40
43.10	17040,0	68,40	83.10	3051,0	12,25
44.20	497,0	2,00	85.10	8012,0	32,16
49.90	276,0	1,11	109.90	310,0	1,24
53.10	677,0	2,72	126.10	24912,0	100,00
56.00	319,0	1,28	127.10	1506,0	6,05
63.00	238,0	0,96			

Spectrometer: HEWLETT PACKARD GC/MS 6890/5973
Inlet system: capillary GC/MS
Scan Range: 40-700 amu
Source temperature: 230 °C

Flow: 1.2 mL/min
Column: DB 17-HT (30 m x 0.25 mm x 0.15 µm)
Programme temperature: 40 °C (3 min); 20 °C/min
(350 °C); 350 °C (20 min)